

Experts in Blood Pressure Management™



PharmaSmart®

ReCharge Instructions

(How to Re-Charge your
Customer's/Patient's/Employee's
Hypertension Management Smart Card)

Turn Over For Instructions

ReCharge Instructions

- 1.) Firmly insert Re-Enrollment Card into the Smart Card Reader. (See Arrow “insert this way”)
- 2.) “**ReCharge** Smart Card” option will be highlighted in Red. Press **GREEN Start Button**.
- 3.) Re-Enrollment Card will debit (1) credit from the card total and eject. This will reoccur every time you **ReCharge** a Smart Card until you have run out of credits.
- 4.) Insert Customer/Patient/Employee Smart Card into Smart Card Reader.
- 5.) Customer/Patient/Employee Smart Card will automatically **ReCharge** and a “**ReCharge** Successful” message will display on the screen.
- 6.) To exit the menu press the **RED Stop Button** until “Exit” is highlighted on the Menu.
- 7.) Press **GREEN Start Button** to Exit Menu.
- 8.) If applicable, tender sale (Reference your SOP or contact your program administrator)

- If **ReCharge** process is unsuccessful (steps 1-8) refer to the following:

A. Reattempt the ReCharge Process: Insert the Customer/Patient/Employee Smart Card again and reattempt steps (2-8). The **ReCharge** process will automatically begin. If the **ReCharge** is unsuccessful the Customer/Patient/Employee Smart Card may become invalid. (If invalid see section B below.)

B. Credit the ReCharge Smart Card: The Re-Enrollment card must be re-inserted into the card reader. The “Credit **ReCharge** Card” process will automatically begin. The Re-Enrollment card will be credited with (1) **ReCharge** and the card will eject.

If after this process is completed, Customer/Patient/Employee Smart Card still reads invalid, or the Re-Enrollment card is not working properly please contact PharmaSmart immediately at (800) 781-0323, or visit us online at www.pharma-smart.com.

***Replenishment:** If your card has run out of credits please contact (800) 781-0323.



Scan this barcode to tender the sale. (*if applicable)

PSI-RCG-041-ME-00